

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111621\
 Data File : VY006761.D
 Acq On : 16 Nov 2021 15:03
 Operator : SY/MD
 Sample : M4684-01
 Misc : 5.31G/5ML/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AU-5-111221

Quant Time: Nov 17 04:07:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	91451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	159164	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	148137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	58765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	71304	59.491	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.980%
35) Dibromofluoromethane	7.728	113	49486	50.617	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.240%
50) Toluene-d8	10.184	98	199467	49.304	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.600%
62) 4-Bromofluorobenzene	12.483	95	61915	45.396	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.800%
Target Compounds					Qvalue	
16) Acetone	3.954	43	3703	10.169	ug/l	90
20) Methylene Chloride	4.734	84	14797	10.941	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111621\
Data File : VY006761.D
Acq On : 16 Nov 2021 15:03
Operator : SY/MD
Sample : M4684-01
Misc : 5.31G/5ML/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AU-5-111221

Quant Time: Nov 17 04:07:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

