

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111622\
 Data File : VY011469.D
 Acq On : 16 Nov 2022 13:32
 Operator : KP/MD
 Sample : N5638-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JC-3-111522

Quant Time: Nov 17 03:15:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	2203	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	3294	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	1846	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	372	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	1146	62.079	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.160%
35) Dibromofluoromethane	7.722	113	688	37.269	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	74.540%
50) Toluene-d8	10.185	98	3433	48.679	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.360%
62) 4-Bromofluorobenzene	12.483	95	401	16.302	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	32.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111622\
Data File : VY011469.D
Acq On : 16 Nov 2022 13:32
Operator : KP/MD
Sample : N5638-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JC-3-111522

Quant Time: Nov 17 03:15:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

