

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111622\
 Data File : VY011481.D
 Acq On : 16 Nov 2022 18:13
 Operator : KP/MD
 Sample : N5641-11
 Misc : 7.18g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
ClientSampleId :
 SOIL-PILE-11

Quant Time: Nov 17 03:17:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.789	168	252304	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	464227	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	435547	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	185201	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	140056	66.244	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.480%
35) Dibromofluoromethane	7.715	113	138316	53.165	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.340%
50) Toluene-d8	10.178	98	540955	54.427	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.860%
62) 4-Bromofluorobenzene	12.483	95	169552	48.910	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.820%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111622\
 Data File : VY011481.D
 Acq On : 16 Nov 2022 18:13
 Operator : KP/MD
 Sample : N5641-11
 Misc : 7.18g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-PILE-11

Quant Time: Nov 17 03:17:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

