

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111622\
 Data File : VY011484.D
 Acq On : 16 Nov 2022 19:23
 Operator : KP/MD
 Sample : N5641-14
 Misc : 6.51g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-PILE-14

Quant Time: Nov 17 03:17:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	253736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	461235	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	446468	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	201468	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	145328	68.350	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.700%
35) Dibromofluoromethane	7.716	113	144810	56.022	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.040%
50) Toluene-d8	10.179	98	547513	55.445	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.880%
62) 4-Bromofluorobenzene	12.483	95	180690	52.461	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111622\
Data File : VY011484.D
Acq On : 16 Nov 2022 19:23
Operator : KP/MD
Sample : N5641-14
Misc : 6.51g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-PILE-14

Quant Time: Nov 17 03:17:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

