

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111623\
 Data File : VY016379.D
 Acq On : 16 Nov 2023 10:38
 Operator : SY/MD
 Sample : VY1116SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1116SBL01

Quant Time: Nov 17 04:01:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	98527	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	175456	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	167205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	73931	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	60202	66.854	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.700%
35) Dibromofluoromethane	7.728	113	55494	55.678	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.360%
50) Toluene-d8	10.185	98	204081	50.492	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.980%
62) 4-Bromofluorobenzene	12.489	95	69733	53.121	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.240%
Target Compounds						
20) Methylene Chloride	4.729	84	10261	6.740	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111623\
Data File : VY016379.D
Acq On : 16 Nov 2023 10:38
Operator : SY/MD
Sample : VY1116SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1116SBL01

Quant Time: Nov 17 04:01:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

