

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111623\
 Data File : VY016392.D
 Acq On : 16 Nov 2023 16:03
 Operator : SY/MD
 Sample : 05418-07
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STONE-3

Quant Time: Nov 17 04:08:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	100507	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.691	114	178957	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.496	117	176728	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	84265	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	63404	69.023	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 138.040%	
35) Dibromofluoromethane	7.716	113	56689	55.764	ug/l	-0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.520%	
50) Toluene-d8	10.185	98	211954	51.414	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.820%	
62) 4-Bromofluorobenzene	12.483	95	76715	57.296	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 114.600%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	17124	13.405	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111623\
Data File : VY016392.D
Acq On : 16 Nov 2023 16:03
Operator : SY/MD
Sample : 05418-07
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
STONE-3

Quant Time: Nov 17 04:08:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

