

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111623\
 Data File : VY016394.D
 Acq On : 16 Nov 2023 16:50
 Operator : SY/MD
 Sample : 05447-04
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP2

Quant Time: Nov 17 04:08:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	98584	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.691	114	173111	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.496	117	167281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	78562	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	61809	68.599	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	137.200%
35) Dibromofluoromethane	7.722	113	55100	56.031	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.060%
50) Toluene-d8	10.185	98	205865	51.624	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.240%
62) 4-Bromofluorobenzene	12.489	95	72278	55.805	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	111.620%
Target Compounds						
20) Methylene Chloride	4.710	84	13553	10.094	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111623\
Data File : VY016394.D
Acq On : 16 Nov 2023 16:50
Operator : SY/MD
Sample : 05447-04
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP2

Quant Time: Nov 17 04:08:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

