

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111623\
 Data File : VY016399.D
 Acq On : 16 Nov 2023 18:48
 Operator : SY/MD
 Sample : 05449-11
 Misc : 5.1g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-SED-04-G

Quant Time: Nov 17 05:19:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	2584	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	4552	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.495	117	4493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	1271	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	1598	67.664	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.320%
35) Dibromofluoromethane	7.722	113	473	18.292	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	36.580%#
50) Toluene-d8	10.185	98	5266	50.219	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.440%
62) 4-Bromofluorobenzene	12.489	95	1515	44.484	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111623\
Data File : VY016399.D
Acq On : 16 Nov 2023 18:48
Operator : SY/MD
Sample : 05449-11
Misc : 5.1g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-SED-04-G

Quant Time: Nov 17 05:19:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

