



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111623\
Data File : VY016377.D
Acq On : 16 Nov 2023 08:28
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Title : SW846 8260
Last Update : Thu Nov 09 00:45:52 2023



AutoFind: Scans 1750, 1751, 1752; Background Corrected with Scan 1742

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	12522	PASS
75	95	30	60	52.1	37592	PASS
95	95	100	100	100.0	72088	PASS
96	95	5	9	7.0	5024	PASS
173	174	0.00	2	1.0	645	PASS
174	95	50	100	88.5	63800	PASS
175	174	5	9	7.1	4518	PASS
176	174	95	101	96.1	61331	PASS
177	176	5	9	6.8	4155	PASS