

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
 Data File : VY006782.D
 Acq On : 17 Nov 2021 12:45
 Operator : SY/MD
 Sample : M4701-01
 Misc : 5.09G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-111521

Quant Time: Nov 18 03:01:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Vimala Arumugam 11/18/2021
 Supervised By :Mahesh Dadoda 11/18/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	2470	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	4405	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	4169	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	1214	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	2380	73.520	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 147.040%			
35) Dibromofluoromethane	7.734	113	1419	52.444	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 104.880%			
50) Toluene-d8	10.185	98	5700	50.908	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 101.820%			
62) 4-Bromofluorobenzene	12.489	95	1512	40.057	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 80.120%			
Target Compounds					Qvalue	
16) Acetone	3.948	43	497	50.531	ug/l	99
20) Methylene Chloride	4.722	84	331m	8.430	ug/l	
85) sec-Butylbenzene	13.251	105	245	2.500	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
Data File : VY006782.D
Acq On : 17 Nov 2021 12:45
Operator : SY/MD
Sample : M4701-01
Misc : 5.09G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-111521

Quant Time: Nov 18 03:01:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Vimala Arumugam 11/18/2021
Supervised By :Mahesh Dadoda 11/18/2021

