

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
 Data File : VY006792.D
 Acq On : 17 Nov 2021 16:42
 Operator : SY/MD
 Sample : M4659-02
 Misc : 9.63G/5ML/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HB-1-(7.5-8)-11-11-21

Quant Time: Nov 18 03:02:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	92944	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	159129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	150445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	93355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	73681	60.486	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.980%	
35) Dibromofluoromethane	7.722	113	50139	51.297	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.600%	
50) Toluene-d8	10.185	98	199862	49.413	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.820%	
62) 4-Bromofluorobenzene	12.483	95	86687	63.573	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 127.140%	
Target Compounds						
						Qvalue
16) Acetone	3.948	43	19093	51.588	ug/l	99
17) Carbon Disulfide	4.204	76	4730	1.489	ug/l	98
25) 2-Butanone	6.996	43	4969	20.309	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
Data File : VY006792.D
Acq On : 17 Nov 2021 16:42
Operator : SY/MD
Sample : M4659-02
Misc : 9.63G/5ML/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HB-1-(7.5-8)-11-11-21

Quant Time: Nov 18 03:02:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

