

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
 Data File : VY006794.D
 Acq On : 17 Nov 2021 18:15
 Operator : SY/MD
 Sample : M4707-09
 Misc : 4.28G/5ML/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 1-COMP

Quant Time: Nov 18 03:03:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	35867	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	58263	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	49165	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	18194	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	24706	52.557	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.120%	
35) Dibromofluoromethane	7.722	113	17841	49.853	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.700%	
50) Toluene-d8	10.185	98	73701	49.767	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.540%	
62) 4-Bromofluorobenzene	12.483	95	20352	40.765	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 81.520%	
Target Compounds						
16) Acetone	3.960	43	2042	14.297	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
Data File : VY006794.D
Acq On : 17 Nov 2021 18:15
Operator : SY/MD
Sample : M4707-09
Misc : 4.28G/5ML/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
1-COMP

Quant Time: Nov 18 03:03:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

