

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
 Data File : VY006796.D
 Acq On : 17 Nov 2021 19:02
 Operator : SY/MD
 Sample : M4707-29
 Misc : 4.75G/5ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 3-COMP

Quant Time: Nov 18 03:03:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	3703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	7091	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.490	117	8425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	2658	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	5155	106.218	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 212.440%#	
35) Dibromofluoromethane	7.722	113	2766	63.505	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 127.000%	
50) Toluene-d8	10.185	98	9002	49.945	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.880%	
62) 4-Bromofluorobenzene	12.483	95	3488	57.404	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 114.800%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.723	84	550	9.742	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
Data File : VY006796.D
Acq On : 17 Nov 2021 19:02
Operator : SY/MD
Sample : M4707-29
Misc : 4.75G/5ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
3-COMP

Quant Time: Nov 18 03:03:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

