

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
 Data File : VY006778.D
 Acq On : 17 Nov 2021 10:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 18 03:00:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.480	0.523	-9.0	99	0.00
3 P	Chloromethane	0.722	0.768	-6.4	95	0.00
4 C	Vinyl Chloride	0.733	0.789	-7.6#	96	0.00
5 T	Bromomethane	0.391	0.464	-18.7	106	0.00
6 T	Chloroethane	0.449	0.494	-10.0	97	0.00
7 T	Trichlorofluoromethane	0.919	1.031	-12.2	102	0.00
8 T	Diethyl Ether	0.306	0.330	-7.8	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.488	0.516	-5.7	96	0.00
10 T	Methyl Iodide	0.610	0.672	-10.2	96	0.00
11 T	Tert butyl alcohol	0.099	0.059	40.4#	61	-0.01
12 CM	1,1-Dichloroethene	0.484	0.535	-10.5#	99	0.00
13 T	Acrolein	0.014	0.015	-7.1	84	0.00
14 T	Allyl chloride	1.090	1.158	-6.2	95	0.00
15 T	Acrylonitrile	0.177	0.178	-0.6	80	0.00
16 T	Acetone	0.199	0.176	11.6	68	0.00
17 T	Carbon Disulfide	1.709	1.816	-6.3	95	0.00
18 T	Methyl Acetate	0.645	0.615	4.7	77	0.00
19 T	Methyl tert-butyl Ether	1.526	1.612	-5.6	88	0.00
20 T	Methylene Chloride	0.704	0.614	12.8	92	0.00
21 T	trans-1,2-Dichloroethene	0.553	0.597	-8.0	96	0.00
22 T	Diisopropyl ether	2.152	2.268	-5.4	92	0.00
23 T	Vinyl Acetate	1.299	1.364	-5.0	86	0.00
24 P	1,1-Dichloroethane	1.098	1.157	-5.4	95	0.00
25 T	2-Butanone	0.267	0.251	6.0	70	0.00
26 T	2,2-Dichloropropane	0.992	1.100	-10.9	100	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.679	-8.6	97	0.00
28 T	Bromochloromethane	0.517	0.492	4.8	86	0.00
29 T	Tetrahydrofuran	0.165	0.160	3.0	74	0.00
30 C	Chloroform	1.056	1.140	-8.0#	97	0.00
31 T	Cyclohexane	1.152	1.188	-3.1	96	0.00
32 T	1,1,1-Trichloroethane	0.973	1.069	-9.9	99	0.00
33 S	1,2-Dichloroethane-d4	0.655	0.618	5.6	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.307	0.281	8.5	83	0.00
36 T	1,1-Dichloropropene	0.543	0.556	-2.4	95	0.00
37 T	Ethyl Acetate	0.363	0.341	6.1	78	0.00
38 T	Carbon Tetrachloride	0.556	0.589	-5.9	98	0.00
39 T	Methylcyclohexane	0.663	0.687	-3.6	96	0.00
40 TM	Benzene	1.530	1.559	-1.9	95	0.00
41 T	Methacrylonitrile	0.305	0.317	-3.9	90	0.00
42 TM	1,2-Dichloroethane	0.510	0.531	-4.1	93	0.00
43 T	Isopropyl Acetate	0.682	0.648	5.0	79	0.00
44 TM	Trichloroethene	0.372	0.412	-10.8	104	0.00
45 C	1,2-Dichloropropane	0.395	0.361	8.6#	85	0.00
46 T	Dibromomethane	0.206	0.205	0.5	88	0.00
47 T	Bromodichloromethane	0.522	0.525	-0.6	92	0.00
48 T	Methyl methacrylate	0.323	0.268	17.0	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
 Data File : VY006778.D
 Acq On : 17 Nov 2021 10:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 18 03:00:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	79	0.00
50 S	Toluene-d8	1.271	1.177	7.4	84	0.00
51 T	4-Methyl-2-Pentanone	0.359	0.299	16.7	66	0.00
52 CM	Toluene	0.929	0.958	-3.1#	96	0.00
53 T	t-1,3-Dichloropropene	0.568	0.575	-1.2	89	0.00
54 T	cis-1,3-Dichloropropene	0.636	0.638	-0.3	92	0.00
55 T	1,1,2-Trichloroethane	0.274	0.271	1.1	91	0.00
56 T	Ethyl methacrylate	0.427	0.428	-0.2	87	0.00
57 T	1,3-Dichloropropane	0.501	0.505	-0.8	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.234	0.206	12.0	76	0.00
59 T	2-Hexanone	0.234	0.242	-3.4	95	0.00
60 T	Dibromochloromethane	0.336	0.338	-0.6	90	0.00
61 T	1,2-Dibromoethane	0.254	0.254	0.0	88	0.00
62 S	4-Bromofluorobenzene	0.428	0.395	7.7	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.365	0.399	-9.3	89	0.00
65 PM	Chlorobenzene	1.017	1.118	-9.9	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.420	-10.5	92	0.00
67 C	Ethyl Benzene	1.928	2.225	-15.4#	98	0.00
68 T	m/p-Xylenes	0.735	0.839	-14.1	96	0.00
69 T	o-Xylene	0.696	0.777	-11.6	94	0.00
70 T	Styrene	1.176	1.342	-14.1	95	0.00
71 P	Bromoform	0.230	0.251	-9.1	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.808	4.277	-12.3	95	0.00
74 T	N-amyl acetate	1.219	1.403	-15.1	79	0.00
75 P	1,1,2,2-Tetrachloroethane	0.704	0.754	-7.1	80	0.00
76 T	1,2,3-Trichloropropane	0.544	0.602	-10.7	83	0.00
77 T	Bromobenzene	0.867	0.957	-10.4	93	0.00
78 T	n-propylbenzene	4.642	5.318	-14.6	95	0.00
79 T	2-Chlorotoluene	2.540	2.901	-14.2	95	0.00
80 T	1,3,5-Trimethylbenzene	3.121	3.557	-14.0	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.274	0.293	-6.9	79	0.00
82 T	4-Chlorotoluene	2.655	3.039	-14.5	94	0.00
83 T	tert-Butylbenzene	2.710	3.105	-14.6	96	0.00
84 T	1,2,4-Trimethylbenzene	3.116	3.483	-11.8	93	0.00
85 T	sec-Butylbenzene	4.036	4.591	-13.8	94	0.00
86 T	p-Isopropyltoluene	3.417	3.857	-12.9	94	0.00
87 T	1,3-Dichlorobenzene	1.727	1.883	-9.0	93	0.00
88 T	1,4-Dichlorobenzene	1.704	1.848	-8.5	93	0.00
89 T	n-Butylbenzene	3.224	3.695	-14.6	93	0.00
90 T	Hexachloroethane	0.616	0.676	-9.7	92	0.00
91 T	1,2-Dichlorobenzene	1.488	1.617	-8.7	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.125	0.128	-2.4	77	0.00
93 T	1,2,4-Trichlorobenzene	0.974	1.036	-6.4	91	0.00
94 T	Hexachlorobutadiene	0.635	0.728	-14.6	101	0.00
95 T	Naphthalene	1.801	2.018	-12.0	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111721\
Data File : VY006778.D
Acq On : 17 Nov 2021 10:54
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Nov 18 03:00:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.811	0.975	-20.2	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6