



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111724\
Data File : VY020316.D
Acq On : 17 Nov 2024 11:14
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111724S.M
Title : SW846 8260
Last Update : Mon Nov 18 03:39:00 2024



AutoFind: Scans 1753, 1754, 1755; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.7	37840	PASS
75	95	30	60	56.0	93304	PASS
95	95	100	100	100.0	166528	PASS
96	95	5	9	6.2	10246	PASS
173	174	0.00	2	1.5	1788	PASS
174	95	50	100	73.8	122845	PASS
175	174	5	9	7.4	9095	PASS
176	174	95	101	95.3	117128	PASS
177	176	5	9	6.7	7870	PASS