

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111821\
 Data File : VY006806.D
 Acq On : 18 Nov 2021 14:20
 Operator : SY/MD
 Sample : M4736-01
 Misc : 5.67G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-202

Quant Time: Nov 19 05:33:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	1331	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	1686	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.489	117	978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	54	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	787	45.115	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.220%
35) Dibromofluoromethane	7.716	113	329	31.769	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	63.540%
50) Toluene-d8	10.185	98	2001	46.693	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.380%
62) 4-Bromofluorobenzene	12.483	95	255	17.650	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	35.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111821\
Data File : VY006806.D
Acq On : 18 Nov 2021 14:20
Operator : SY/MD
Sample : M4736-01
Misc : 5.67G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-202

Quant Time: Nov 19 05:33:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

