

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111821\
 Data File : VY006808.D
 Acq On : 18 Nov 2021 15:08
 Operator : SY/MD
 Sample : M4707-29
 Misc : 4.15G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 3-COMP

Quant Time: Nov 19 05:34:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	115766	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	200008	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	178991	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	66284	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	86138	56.772	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.540%
35) Dibromofluoromethane	7.728	113	63943	52.048	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.100%
50) Toluene-d8	10.179	98	252928	49.752	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.500%
62) 4-Bromofluorobenzene	12.483	95	73833	43.080	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111821\
Data File : VY006808.D
Acq On : 18 Nov 2021 15:08
Operator : SY/MD
Sample : M4707-29
Misc : 4.15G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
3-COMP

Quant Time: Nov 19 05:34:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

