

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111821\
 Data File : VY006813.D
 Acq On : 18 Nov 2021 17:44
 Operator : SY/MD
 Sample : M4766-07
 Misc : 5.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-27165

Quant Time: Nov 19 05:35:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	115002	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	200589	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	188497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	76915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	88338	58.609	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163		Recovery	= 117.220%	
35) Dibromofluoromethane	7.722	113	63548	51.577	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147		Recovery	= 103.160%	
50) Toluene-d8	10.185	98	262711	51.526	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140		Recovery	= 103.060%	
62) 4-Bromofluorobenzene	12.483	95	86291	50.203	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144		Recovery	= 100.400%	
Target Compounds						
16) Acetone	3.954	43	4362	9.525	ug/l #	87
20) Methylene Chloride	4.728	84	12627	6.242	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111821\
Data File : VY006813.D
Acq On : 18 Nov 2021 17:44
Operator : SY/MD
Sample : M4766-07
Misc : 5.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-27165

Quant Time: Nov 19 05:35:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

