



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111821\
 Data File : VY006801.D
 Acq On : 18 Nov 2021 09:33
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Title : SW846 8260
 Last Update : Tue Nov 16 04:20:03 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.4	20915	PASS
75	95	30	60	54.9	53765	PASS
95	95	100	100	100.0	97848	PASS
96	95	5	9	7.0	6842	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.7	84861	PASS
175	174	5	9	8.2	6972	PASS
176	174	95	101	97.7	82883	PASS
177	176	5	9	8.0	6646	PASS