

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
 Data File : VY006823.D
 Acq On : 19 Nov 2021 12:24
 Operator : SY/MD
 Sample : M4766-01RE
 Misc : 5.09G/5ML/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-20058RE

Quant Time: Nov 22 03:18:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	27901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	42591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	31202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	9029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	16643	45.513	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.020%
35) Dibromofluoromethane	7.722	113	13143	50.239	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.480%
50) Toluene-d8	10.185	98	51532	47.601	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.200%
62) 4-Bromofluorobenzene	12.483	95	10994	30.124	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	60.240%
Target Compounds						Qvalue
20) Methylene Chloride	4.722	84	2171	3.351	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
Data File : VY006823.D
Acq On : 19 Nov 2021 12:24
Operator : SY/MD
Sample : M4766-01RE
Misc : 5.09G/5ML/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-20058RE

Quant Time: Nov 22 03:18:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

