

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
 Data File : VY006824.D
 Acq On : 19 Nov 2021 12:47
 Operator : SY/MD
 Sample : M4766-03RE
 Misc : 5.06G/5ML/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-26706

Quant Time: Nov 22 03:19:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	100492	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	173291	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	156669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	58372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	74756	56.759	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.520%
35) Dibromofluoromethane	7.728	113	54796	51.480	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.960%
50) Toluene-d8	10.185	98	221115	50.200	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.400%
62) 4-Bromofluorobenzene	12.483	95	62925	42.376	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.760%
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	11222	6.411	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
Data File : VY006824.D
Acq On : 19 Nov 2021 12:47
Operator : SY/MD
Sample : M4766-03RE
Misc : 5.06G/5ML/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-26706

Quant Time: Nov 22 03:19:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

