

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
 Data File : VY006825.D
 Acq On : 19 Nov 2021 13:11
 Operator : SY/MD
 Sample : M4758-01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP1

Quant Time: Nov 22 03:19:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	21723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	32180	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	25897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	8710	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.154	65	12908	45.338	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.680%
35) Dibromofluoromethane	7.721	113	9894	50.055	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.120%
50) Toluene-d8	10.184	98	40018	48.925	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.840%
62) 4-Bromofluorobenzene	12.483	95	10385	37.661	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.320%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.716	84	2669	7.423	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
Data File : VY006825.D
Acq On : 19 Nov 2021 13:11
Operator : SY/MD
Sample : M4758-01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP1

Quant Time: Nov 22 03:19:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

