

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
 Data File : VY006826.D
 Acq On : 19 Nov 2021 13:35
 Operator : SY/MD
 Sample : M4577-03
 Misc : 6.45G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-3

Quant Time: Nov 19 18:16:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	101750	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	172897	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	154838	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	60509	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	78102	58.567	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.140%
35) Dibromofluoromethane	7.722	113	54173	51.010	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.020%
50) Toluene-d8	10.185	98	221614	50.428	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.860%
62) 4-Bromofluorobenzene	12.483	95	63251	42.692	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
Data File : VY006826.D
Acq On : 19 Nov 2021 13:35
Operator : SY/MD
Sample : M4577-03
Misc : 6.45G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-3

Quant Time: Nov 19 18:16:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

