

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
 Data File : VY006836.D
 Acq On : 19 Nov 2021 17:31
 Operator : SY/MD
 Sample : M4758-25
 Misc : 5.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP13

Quant Time: Nov 22 03:20:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	22135	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	31295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	23445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	7712	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	12642	43.577	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.160%
35) Dibromofluoromethane	7.722	113	9626	50.076	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.160%
50) Toluene-d8	10.185	98	38746	48.709	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.420%
62) 4-Bromofluorobenzene	12.483	95	8463	31.559	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	63.120%
Target Compounds						
16) Acetone	3.961	43	1110	12.593	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
Data File : VY006836.D
Acq On : 19 Nov 2021 17:31
Operator : SY/MD
Sample : M4758-25
Misc : 5.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP13

Quant Time: Nov 22 03:20:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

