

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
 Data File : VY006842.D
 Acq On : 19 Nov 2021 20:16
 Operator : SY/MD
 Sample : M4758-11
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP6

Quant Time: Nov 22 03:21:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	56228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	90660	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	76265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	25618	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	32901	44.646	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.300%
35) Dibromofluoromethane	7.728	113	27087	48.642	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.280%
50) Toluene-d8	10.185	98	113511	49.258	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.520%
62) 4-Bromofluorobenzene	12.483	95	26852	34.565	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	69.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111921\
Data File : VY006842.D
Acq On : 19 Nov 2021 20:16
Operator : SY/MD
Sample : M4758-11
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP6

Quant Time: Nov 22 03:21:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

