

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY112020\
 Data File : VY003564.D
 Acq On : 20 Nov 2020 13:57
 Operator : SY/MD
 Sample : VY1120SBL01
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1120SBL01

Quant Time: Nov 21 02:20:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112020S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:00:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	121260	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	254347	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	229789	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	99636	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.14	65	83648	66.44	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.88%
35) Dibromofluoromethane	7.72	113	83232	49.78	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.56%
50) Toluene-d8	10.18	98	299527	46.07	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.14%
62) 4-Bromofluorobenzene	12.48	95	92394	44.78	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.56%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.71	84	14895	5.962	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY112020\
Data File : VY003564.D
Acq On : 20 Nov 2020 13:57
Operator : SY/MD
Sample : VY1120SBL01
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1120SBL01

Quant Time: Nov 21 02:20:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112020S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:00:42 2020
Response via : Initial Calibration

