

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY112020\
 Data File : VY003574.D
 Acq On : 20 Nov 2020 17:40
 Operator : SY/MD
 Sample : L4836-06
 Misc : 5.8G/5ML/MSVOA Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-06

Quant Time: Nov 21 03:05:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112020S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:00:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.79	168	94140	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	219615	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	210904	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	97008	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.14	65	76445	78.21	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	156.42%
35) Dibromofluoromethane	7.72	113	77362	53.59	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.18%
50) Toluene-d8	10.17	98	270788	48.24	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.48%
62) 4-Bromofluorobenzene	12.48	95	85177	47.81	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.62%

Target Compounds

					Qvalue	
16) Acetone	3.98	43	10421	31.930	ug/l	91
95) Naphthalene	15.23	128	3283	1.328	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY112020\
Data File : VY003574.D
Acq On : 20 Nov 2020 17:40
Operator : SY/MD
Sample : L4836-06
Misc : 5.8G/5ML/MSVOA Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-06

Quant Time: Nov 21 03:05:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112020S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:00:42 2020
Response via : Initial Calibration

