

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112122\
 Data File : VY011525.D
 Acq On : 21 Nov 2022 15:39
 Operator : KP/MD
 Sample : N5678-01RE
 Misc : 6.85g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 250FB-BP-BACKFILL-01RE

Quant Time: Nov 22 01:46:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	178928	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	347041	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	332193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	149028	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	129970	71.537	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.080%
35) Dibromofluoromethane	7.722	113	114269	53.488	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.980%
50) Toluene-d8	10.179	98	427031	55.006	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.020%
62) 4-Bromofluorobenzene	12.483	95	138277	48.174	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112122\
Data File : VY011525.D
Acq On : 21 Nov 2022 15:39
Operator : KP/MD
Sample : N5678-01RE
Misc : 6.85g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
250FB-BP-BACKFILL-01RE

Quant Time: Nov 22 01:46:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

