

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112124\
 Data File : VY020374.D
 Acq On : 21 Nov 2024 09:59
 Operator : SY/MD
 Sample : VY1121SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1121SBL01

Quant Time: Nov 21 13:56:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	139506	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	270751	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	245370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	89046	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	83114	51.857	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.720%
35) Dibromofluoromethane	7.640	113	84115	48.450	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.900%
50) Toluene-d8	10.109	98	329763	48.218	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.440%
62) 4-Bromofluorobenzene	12.408	95	104043	47.324	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	94.640%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112124\
Data File : VY020374.D
Acq On : 21 Nov 2024 09:59
Operator : SY/MD
Sample : VY1121SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1121SBL01

Quant Time: Nov 21 13:56:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

