

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112124\
 Data File : VY020379.D
 Acq On : 21 Nov 2024 12:12
 Operator : SY/MD
 Sample : P4852-03
 Misc : 5.14g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-3

Quant Time: Nov 21 13:57:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	134247	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	257659	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	237210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	92815	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	82024	53.181	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	106.360%	
35) Dibromofluoromethane	7.640	113	81524	49.343	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	98.680%	
50) Toluene-d8	10.109	98	316937	48.697	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	97.400%	
62) 4-Bromofluorobenzene	12.408	95	104261	49.833	ug/l	0.00
Spiked Amount	50.000	Range 29 - 146	Recovery	=	99.660%	
Target Compounds						
16) Acetone	3.867	43	2702	7.308	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112124\
Data File : VY020379.D
Acq On : 21 Nov 2024 12:12
Operator : SY/MD
Sample : P4852-03
Misc : 5.14g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-3

Quant Time: Nov 21 13:57:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

