

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112124\
 Data File : VY020394.D
 Acq On : 21 Nov 2024 18:04
 Operator : SY/MD
 Sample : P4925-07
 Misc : 5.53g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-758-VOC

Quant Time: Nov 22 00:28:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	124823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	228431	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	201781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	77403	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	72776	50.748	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.500%
35) Dibromofluoromethane	7.640	113	69838	47.679	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.360%
50) Toluene-d8	10.109	98	274111	47.506	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.020%
62) 4-Bromofluorobenzene	12.407	95	87260	47.044	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	94.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112124\
Data File : VY020394.D
Acq On : 21 Nov 2024 18:04
Operator : SY/MD
Sample : P4925-07
Misc : 5.53g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-758-VOC

Quant Time: Nov 22 00:28:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

