

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112124\
 Data File : VY020396.D
 Acq On : 21 Nov 2024 18:50
 Operator : SY/MD
 Sample : P4938-07
 Misc : 5.65g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-734-VOC

Quant Time: Nov 22 00:29:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	153170	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	294641	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	279410	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	116711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	94414	53.652	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.300%
35) Dibromofluoromethane	7.634	113	93869	49.684	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.360%
50) Toluene-d8	10.109	98	361444	48.565	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.140%
62) 4-Bromofluorobenzene	12.408	95	124910	52.209	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	104.420%
Target Compounds						
16) Acetone	3.873	43	5729	13.580	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112124\
Data File : VY020396.D
Acq On : 21 Nov 2024 18:50
Operator : SY/MD
Sample : P4938-07
Misc : 5.65g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-734-VOC

Quant Time: Nov 22 00:29:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

