



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112124\
Data File : VY020372.D
Acq On : 21 Nov 2024 08:32
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Title : SW846 8260
Last Update : Wed Nov 20 04:38:24 2024



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1745

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	18537	PASS
75	95	30	60	55.9	48139	PASS
95	95	100	100	100.0	86101	PASS
96	95	5	9	6.7	5771	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	73.7	63419	PASS
175	174	5	9	7.9	5005	PASS
176	174	95	101	95.6	60600	PASS
177	176	5	9	6.7	4051	PASS