

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112221\
 Data File : VY006851.D
 Acq On : 22 Nov 2021 14:19
 Operator : SY/MD
 Sample : M4801-01
 Misc : 5.59G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SITE-1-DRILL-CUTTINGS-20211119

Quant Time: Nov 23 04:47:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	88044	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	149269	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	144232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	57353	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	67214	58.248	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.500%
35) Dibromofluoromethane	7.722	113	47309	51.598	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.200%
50) Toluene-d8	10.185	98	190306	50.158	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.320%
62) 4-Bromofluorobenzene	12.483	95	57817	45.202	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112221\
Data File : VY006851.D
Acq On : 22 Nov 2021 14:19
Operator : SY/MD
Sample : M4801-01
Misc : 5.59G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SITE-1-DRILL-CUTTINGS-20211119

Quant Time: Nov 23 04:47:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

