

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112222\
 Data File : VY011545.D
 Acq On : 22 Nov 2022 14:04
 Operator : KP/MD
 Sample : N5749-01
 Misc : 5.60g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-82

Quant Time: Nov 23 04:07:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	123275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	231186	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	214841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	88338	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	84370	67.403	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.800%
35) Dibromofluoromethane	7.716	113	75754	53.229	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.460%
50) Toluene-d8	10.179	98	283844	54.885	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.760%
62) 4-Bromofluorobenzene	12.483	95	85643	44.790	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112222\
Data File : VY011545.D
Acq On : 22 Nov 2022 14:04
Operator : KP/MD
Sample : N5749-01
Misc : 5.60g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-82

Quant Time: Nov 23 04:07:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

