

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112222\
 Data File : VY011550.D
 Acq On : 22 Nov 2022 16:01
 Operator : KP/MD
 Sample : N5753-02
 Misc : 6.24g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11986

Quant Time: Nov 23 04:08:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	190919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	362081	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	332020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	140505	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	128276	66.170	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.340%
35) Dibromofluoromethane	7.716	113	115163	51.667	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.340%
50) Toluene-d8	10.179	98	442893	54.680	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.360%
62) 4-Bromofluorobenzene	12.483	95	133852	44.696	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112222\
Data File : VY011550.D
Acq On : 22 Nov 2022 16:01
Operator : KP/MD
Sample : N5753-02
Misc : 6.24g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11986

Quant Time: Nov 23 04:08:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

