


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112222\
Data File : VY011537.D
Acq On : 22 Nov 2022 09:18
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Title : SW846 8260
Last Update : Sat Nov 19 22:37:14 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.0	15989	PASS
75	95	30	60	54.0	43128	PASS
95	95	100	100	100.0	79912	PASS
96	95	5	9	6.7	5380	PASS
173	174	0.00	2	1.2	802	PASS
174	95	50	100	81.6	65173	PASS
175	174	5	9	7.6	4982	PASS
176	174	95	101	96.7	63035	PASS
177	176	5	9	7.2	4544	PASS