

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112321\
 Data File : VY006865.D
 Acq On : 23 Nov 2021 13:06
 Operator : SY/MD
 Sample : M4758-29
 Misc : 5.04G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP15

Quant Time: Nov 24 02:04:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	91361	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	151859	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	143099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	51721	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	65557	54.750	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.500%
35) Dibromofluoromethane	7.728	113	46811	50.185	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.360%
50) Toluene-d8	10.185	98	194580	50.410	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.820%
62) 4-Bromofluorobenzene	12.483	95	53259	40.928	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112321\
Data File : VY006865.D
Acq On : 23 Nov 2021 13:06
Operator : SY/MD
Sample : M4758-29
Misc : 5.04G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP15

Quant Time: Nov 24 02:04:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

