

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112321\
 Data File : VY006866.D
 Acq On : 23 Nov 2021 13:30
 Operator : SY/MD
 Sample : M4758-31RE
 Misc : 5.06G/5ML/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP16RE

Quant Time: Nov 24 02:04:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	25463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	41230	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	41997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	17417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	21513	64.464	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.920%
35) Dibromofluoromethane	7.728	113	13374	52.809	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.620%
50) Toluene-d8	10.185	98	53566	51.113	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.220%
62) 4-Bromofluorobenzene	12.483	95	15659	44.322	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.640%
Target Compounds						Qvalue
16) Acetone	3.966	43	1373	13.541	ug/l #	87
20) Methylene Chloride	4.735	84	2922	6.690	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112321\
Data File : VY006866.D
Acq On : 23 Nov 2021 13:30
Operator : SY/MD
Sample : M4758-31RE
Misc : 5.06G/5ML/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP16RE

Quant Time: Nov 24 02:04:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

