

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112322\
 Data File : VY011583.D
 Acq On : 23 Nov 2022 18:18
 Operator : KP/MD
 Sample : N5780-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-M

Quant Time: Nov 24 00:39:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	173790	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	326992	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	320288	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	145482	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	124944	70.804	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.600%
35) Dibromofluoromethane	7.716	113	109044	54.171	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.340%
50) Toluene-d8	10.179	98	405805	55.477	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.960%
62) 4-Bromofluorobenzene	12.483	95	133161	49.237	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112322\
Data File : VY011583.D
Acq On : 23 Nov 2022 18:18
Operator : KP/MD
Sample : N5780-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-M

Quant Time: Nov 24 00:39:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

