

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY112420\
 Data File : VY003599.D
 Acq On : 24 Nov 2020 13:05
 Operator : SY/MD
 Sample : L4860-05
 Misc : 5.56G/10ML/MSVOA Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 C0B20

Quant Time: Nov 25 06:17:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\SOM2YLM111920S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 25 06:01:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.69	114	184985	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.49	117	164874	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.42	152	67751	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.25	65	61170	30.00	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	120.00%
7) Chloroethane-d5	2.77	69	48544	25.66	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.64%
10) 1,1-Dichloroethene-d2	3.85	63	100034	19.85	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.40%
20) 2-Butanone-d5	6.91	46	38178	37.14	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.28%
24) Chloroform-d	7.49	84	125879	26.77	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	107.08%
26) 1,2-Dichloroethane-d4	8.14	65	67708	25.69	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.76%
29) Benzene-d6	8.11	84	266371	27.73	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	110.92%
33) 1,2-Dichloropropane-d6	9.13	67	79935	28.38	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	113.52%
37) Toluene-d8	10.18	98	234162	27.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	108.84%
38) trans-1,3-Dichloropropene-	10.44	79	34927	24.85	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.40%
39) 2-Hexanone-d5	10.79	63	26478	36.15	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.30%
48) 1,1,2,2-Tetrachloroethane-	12.56	84	51762	21.10	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.40%
61) 1,2-Dichlorobenzene-d4	13.72	152	60629	27.72	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	110.88%

Target Compounds					Ovalue
16) Methylene chloride	4.72	84	6390	2.015 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY112420\
Data File : VY003599.D
Acq On : 24 Nov 2020 13:05
Operator : SY/MD
Sample : L4860-05
Misc : 5.56G/10ML/MSVOA Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
C0B20

Quant Time: Nov 25 06:17:54 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\SOM2YLM111920S.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 25 06:01:46 2020
Response via : Initial Calibration

