

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY112420\
 Data File : VY003610.D
 Acq On : 24 Nov 2020 17:02
 Operator : SY/MD
 Sample : L4861-01
 Misc : 5.22G/10ML/MSVOA Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 C0B23

Quant Time: Nov 25 06:45:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\SOM2YLM111920S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 25 06:01:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.69	114	164690	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.49	117	149046	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.42	152	61152	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.25	65	48765	26.87	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	107.48%
7) Chloroethane-d5	2.77	69	39169	23.26	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.04%
10) 1,1-Dichloroethene-d2	3.86	63	80317	17.90	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.60%
20) 2-Butanone-d5	6.90	46	26396	28.84	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	57.68%
24) Chloroform-d	7.49	84	104358	24.93	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.72%
26) 1,2-Dichloroethane-d4	8.14	65	54853	23.38	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.52%
29) Benzene-d6	8.11	84	214958	24.75	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.00%
33) 1,2-Dichloropropane-d6	9.13	67	64694	25.41	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.64%
37) Toluene-d8	10.18	98	186285	23.94	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.76%
38) trans-1,3-Dichloropropene-	10.44	79	26953	21.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.84%
39) 2-Hexanone-d5	10.79	63	18553	28.02	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	56.04%
48) 1,1,2,2-Tetrachloroethane-	12.56	84	39488	17.80	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	71.20%
61) 1,2-Dichlorobenzene-d4	13.72	152	48719	24.68	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.72%

Target Compounds					Ovalue
16) Methylene chloride	4.72	84	7841	2.777 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY112420\
Data File : VY003610.D
Acq On : 24 Nov 2020 17:02
Operator : SY/MD
Sample : L4861-01
Misc : 5.22G/10ML/MSVOA Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
C0B23

Quant Time: Nov 25 06:45:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\SOM2YLM111920S.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 25 06:01:46 2020
Response via : Initial Calibration

