

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112421\
 Data File : VY006884.D
 Acq On : 24 Nov 2021 18:02
 Operator : SY/MD
 Sample : M4786-02
 Misc : 5.03G/5ML/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-279

Quant Time: Nov 26 09:58:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	88885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	148004	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	131130	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	56405	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	58779	50.456	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.920%
35) Dibromofluoromethane	7.728	113	42683	46.951	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.900%
50) Toluene-d8	10.185	98	185919	49.421	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.489	95	54486	42.962	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112421\
Data File : VY006884.D
Acq On : 24 Nov 2021 18:02
Operator : SY/MD
Sample : M4786-02
Misc : 5.03G/5ML/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-279

Quant Time: Nov 26 09:58:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

