

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112423\
 Data File : VY016475.D
 Acq On : 24 Nov 2023 10:42
 Operator : SY/MD
 Sample : VY1124SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1124SBL01

Quant Time: Nov 27 04:42:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	102655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	178793	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	173069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	76312	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	58776	62.646	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	125.300%	
35) Dibromofluoromethane	7.722	113	56751	55.876	ug/l	-0.01
Spiked Amount	50.000	Range 54 - 147	Recovery	=	111.760%	
50) Toluene-d8	10.185	98	207314	50.335	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	100.660%	
62) 4-Bromofluorobenzene	12.489	95	71354	53.341	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	106.680%	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.722	84	8595	4.685	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112423\
Data File : VY016475.D
Acq On : 24 Nov 2023 10:42
Operator : SY/MD
Sample : VY1124SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1124SBL01

Quant Time: Nov 27 04:42:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

