

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112423\
 Data File : VY016487.D
 Acq On : 24 Nov 2023 16:15
 Operator : SY/MD
 Sample : 05519-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-02-112123

Quant Time: Nov 27 04:47:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	14360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	22157	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	17967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	7304	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	6850	52.193	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.380%
35) Dibromofluoromethane	7.728	113	7178	57.029	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.060%
50) Toluene-d8	10.191	98	24272	47.554	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.100%
62) 4-Bromofluorobenzene	12.495	95	6445	38.878	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.760%
Target Compounds						
16) Acetone	3.960	43	1663	76.602	ug/l	94
20) Methylene Chloride	4.735	84	1085	3.863	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112423\
Data File : VY016487.D
Acq On : 24 Nov 2023 16:15
Operator : SY/MD
Sample : 05519-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-02-112123

Quant Time: Nov 27 04:47:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

