

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112423\
 Data File : VY016489.D
 Acq On : 24 Nov 2023 17:02
 Operator : SY/MD
 Sample : 05434-04
 Misc : 5.43g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-2

Quant Time: Nov 27 04:47:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	95767	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	167817	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	161034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	65200	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	55962	63.937	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.880%
35) Dibromofluoromethane	7.734	113	54266	56.924	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.840%
50) Toluene-d8	10.191	98	197550	51.101	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.200%
62) 4-Bromofluorobenzene	12.495	95	62670	49.913	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.820%
Target Compounds						
20) Methylene Chloride	4.729	84	10067	6.838	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112423\
Data File : VY016489.D
Acq On : 24 Nov 2023 17:02
Operator : SY/MD
Sample : 05434-04
Misc : 5.43g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-2

Quant Time: Nov 27 04:47:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

