

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112423\
 Data File : VY016493.D
 Acq On : 24 Nov 2023 18:36
 Operator : SY/MD
 Sample : 05434-12
 Misc : 6.52g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-6

Quant Time: Nov 27 04:49:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	96658	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.709	114	166108	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.507	117	160785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	69828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.160	65	55969	63.355	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	126.720%	
35) Dibromofluoromethane	7.740	113	53640	56.846	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	113.700%	
50) Toluene-d8	10.197	98	194275	50.771	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	101.540%	
62) 4-Bromofluorobenzene	12.501	95	64691	52.053	ug/l	0.01
Spiked Amount	50.000	Range 30 - 143	Recovery	=	104.100%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.746	84	12444	9.215	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112423\
Data File : VY016493.D
Acq On : 24 Nov 2023 18:36
Operator : SY/MD
Sample : 05434-12
Misc : 6.52g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-6

Quant Time: Nov 27 04:49:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

