

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112423\
 Data File : VY016494.D
 Acq On : 24 Nov 2023 18:59
 Operator : SY/MD
 Sample : 05434-14
 Misc : 6.65g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-7

Quant Time: Nov 27 04:49:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	98218	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.715	114	173473	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.508	117	165013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	72604	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	56951	63.443	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.880%	
35) Dibromofluoromethane	7.746	113	54703	55.511	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.020%	
50) Toluene-d8	10.197	98	201917	50.528	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.060%	
62) 4-Bromofluorobenzene	12.501	95	66748	51.428	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	= 102.860%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.747	84	7063	3.496	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112423\
Data File : VY016494.D
Acq On : 24 Nov 2023 18:59
Operator : SY/MD
Sample : 05434-14
Misc : 6.65g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-7

Quant Time: Nov 27 04:49:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

