

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112425\
 Data File : VY023791.D
 Acq On : 24 Nov 2025 15:34
 Operator : SY/MD
 Sample : Q3710-03
 Misc : 4.70g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMPRESSOR-EXCAVATION

Quant Time: Nov 26 02:12:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 05 04:41:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	697466	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	1288188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1262537	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	508090	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	386658	64.990	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.980%
35) Dibromofluoromethane	7.634	113	417905	51.912	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.820%
50) Toluene-d8	10.109	98	1697083	56.592	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	113.180%
62) 4-Bromofluorobenzene	12.407	95	498583	51.412	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	102.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112425\
Data File : VY023791.D
Acq On : 24 Nov 2025 15:34
Operator : SY/MD
Sample : Q3710-03
Misc : 4.70g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMPRESSOR-EXCAVATION

Quant Time: Nov 26 02:12:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 05 04:41:50 2025
Response via : Initial Calibration

