

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112425\
 Data File : VY023795.D
 Acq On : 24 Nov 2025 17:12
 Operator : SY/MD
 Sample : Q3707-01
 Misc : 5.47g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 V933

Quant Time: Nov 26 02:14:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 05 04:41:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	472933	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	857142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	835519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	316923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	259410	64.303	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.600%
35) Dibromofluoromethane	7.634	113	218507	40.792	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.580%
50) Toluene-d8	10.109	98	991728	49.702	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.400%
62) 4-Bromofluorobenzene	12.401	95	318272	49.323	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.640%
Target Compounds						
16) Acetone	3.872	43	9255	9.229	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112425\
Data File : VY023795.D
Acq On : 24 Nov 2025 17:12
Operator : SY/MD
Sample : Q3707-01
Misc : 5.47g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
V933

Quant Time: Nov 26 02:14:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 05 04:41:50 2025
Response via : Initial Calibration

